Annex to declaration of accreditation (scope of accreditation)

Normative document: EN ISO/IEC 17025:2017

Registration number: K 162

of Element Materials Technology Rotterdam B.V. Calibration Laboratory

This annex is valid from: 20-06-2024 to 01-09-2028 Replaces annex dated: 08-11-2023

Location(s) where activities are performed under accreditation

Head Office Voorerf 18 4824 GN Breda The Netherlands

Location	Abbreviation/ location code	
Zekeringstraat 33 1014 BV Amsterdam Nederland	AM	

HCS code	Measured quantity, Instrument, Measure	Range	CMC ¹	Remarks	Location
FQ 0 0	Force				
FQ 0 0	Load cells	Tension 20 N – 600 kN	2,5·10 ⁻³ · <i>F</i>		AM
		Compression 20 N – 600 kN	4·10 ⁻³ · <i>F</i>		

This annex has been approved by the Board of the Dutch Accreditation Council, on its behalf,

J.A.W.M. de Haas

Dutch Accreditation Council RvA Pagina 1 van 1

¹ Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty, *U*, is calculated according to EA-4/02 "Evaluation of the Uncertainty of Measurement in Calibration".