

Annex to declaration of accreditation (scope of accreditation)

Normative document: EN ISO/IEC 17025:2017

Registration number: **K 162**

of **Element Materials Technology Rotterdam B.V.**  
**Calibration Laboratory**

This annex is valid from: **20-06-2024** to **01-09-2028**

Replaces annex dated: **08-11-2023**

**Location(s) where activities are performed under accreditation**

**Head Office**

Voorerf 18  
4824 GN  
Breda  
The Netherlands

Location	Abbreviation/ location code
Zekeringstraat 33 1014 BV Amsterdam Nederland	AM

HCS code	Measured quantity, Instrument, Measure	Range	CMC <sup>1</sup>	Remarks	Location
FQ 0 0	Force				
FQ 0 0	Load cells	Tension 20 N – 600 kN	$2,5 \cdot 10^{-3} \cdot F$		AM
		Compression 20 N – 600 kN	$4 \cdot 10^{-3} \cdot F$		

This annex has been approved by the Board of the  
Dutch Accreditation Council, on its behalf,

J.A.W.M. de Haas

<sup>1</sup> Calibration and Measurement Capability (CMC): Demonstrated measurement uncertainty, with coverage probability of 95%, in a given measurement point or measurement range. Measurement uncertainty,  $U$ , is calculated according to EA-4/02 "Evaluation of the Uncertainty of Measurement in Calibration".